



Research Article

DEVELOPMENT OF A CLOUD-CONNECTED IOT INFRASTRUCTURE FOR INTELLIGENT PATIENT HEALTH ASSESSMENT

^{1*}Pugazh Gowrishankar, ²Ashwini L, ³Vijai Krishna V, ⁴Madhumitha N and ⁵Geetha C

^{1*}PERI Institute of Technology, Chennai- 48, Tamil Nadu, India

²PERI College of Arts and Science, Chennai - 48, Tamil Nadu, India

³PERI College of Physiotherapy, Chennai - 48, Tamil Nadu, India

⁴PERI College of Pharmacy, Chennai - 48, Tamil Nadu, India

⁵PERI College of Nursing, Chennai - 48, Tamil Nadu, India

Article History: Received 26th September 2025; Accepted 26th November 2025; Published 1st December 2025

ABSTRACT

The rapid evolution of Internet of Things (IoT) technologies has significantly transformed modern healthcare systems by enabling real-time physiological monitoring, seamless data transmission, and intelligent clinical decision support. This study presents the development of a cloud-connected IoT infrastructure designed for continuous patient health assessment using multi-sensor data acquisition. The proposed system integrates temperature, blood pressure, and heart-rate sensors with an AT89S52 microcontroller to collect vital signs, which are transmitted through wireless communication to a cloud server for storage, analysis, and visualization. The cloud platform enables remote access for healthcare professionals, allowing timely diagnosis, proactive intervention, and comprehensive patient monitoring. Experimental results demonstrate the system's reliability, accuracy, and ability to provide health status updates every 60 seconds. This work highlights the potential of IoT-enabled healthcare systems to enhance patient outcomes, reduce hospital workload, and support scalable telemedicine applications. The study also outlines future enhancements including AI-driven analytics, predictive diagnostics, and extended multi-parameter sensing.

Keywords: IoT-Based Healthcare, Cloud Computing, Patient Monitoring, Biomedical Sensors, AT89S52.

INTRODUCTION

The increasing burden on global healthcare systems, coupled with the rise in chronic diseases and aging populations, has intensified the demand for scalable and intelligent patient monitoring solutions. Conventional healthcare frameworks rely heavily on periodic clinical visits and manual assessment, which often results in delayed diagnosis and limited real-time insight into patients' health conditions. To address these limitations, the Internet of Things (IoT) has emerged as a transformative paradigm capable of interconnecting medical devices, sensors, and cloud platforms to create intelligent, automated, and patient-centered healthcare environments. IoT-enabled healthcare systems facilitate continuous monitoring through wearable, implantable, or ambient sensors that can capture critical physiological parameters

such as body temperature, blood pressure, and heart rate. These systems further ensure seamless communication between sensing nodes and cloud servers using wireless sensor networks (WSNs). By leveraging cloud computing, the collected data can be securely stored, processed, and accessed in real time by healthcare professionals from any location. This enhances clinical decision-making, enables early detection of abnormalities, and promotes efficient health management. The concept of cloud-connected IoT infrastructures specifically addresses challenges such as data accessibility, scalability, interoperability, and remote supervision. In this context, body area networks (BANs) and wireless body area networks (WBANs) have become essential components, enabling short-range communication around the human body. Their integration with cloud platforms ensures robust medical monitoring without

*Corresponding Author: Pugazh Gowrishankar, PERI Institute of Technology, Chennai- 48, Tamil Nadu, India Email: publications@peri.ac.in

constant physical presence of medical personnel. This paper presents the development of an intelligent patient health assessment system that employs an AT89S52 microcontroller to interface with biomedical sensors. The system collects real-time vital signs and transmits the data to a cloud server for continuous tracking and automated evaluation. The study aims to demonstrate the technical feasibility, efficiency, and reliability of cloud-enabled IoT systems in healthcare settings. Additionally, the paper discusses potential enhancements including AI-powered analytics, predictive health modeling, and integration with advanced medical services.

The integration of IoT into healthcare has revolutionized medical monitoring by enabling pervasive sensing, remote supervision, and automated clinical workflows. IoT systems enhance the accessibility, accuracy, and reliability of patient data, thereby supporting continuous health assessment (Atzori *et al.*, 2010). Recent studies emphasize that IoT-healthcare models reduce hospital workloads and improve chronic disease management through uninterrupted data acquisition and cloud-assisted analytics

(Borgia, 2014). These advancements prove crucial for early diagnosis and timely interventions. Wireless Sensor Networks (WSNs) constitute the backbone of IoT-enabled health monitoring. BANs and WBANs are essential for short-range communication between biomedical sensors and processing units. They ensure energy efficiency, mobility support, and reliable data transmission within a limited range around the human body (Delmastro, 2012). WBANs significantly improve patient comfort and allow seamless integration of wearable or implantable sensors for real-time health tracking (Tartarisco *et al.*, 2012). Biomedical sensors play a significant role in capturing physiological parameters such as blood pressure, heart rate, and temperature. Accurate sensing of these parameters is essential for diagnosing cardiovascular conditions, detecting fever, and understanding patient well-being trends (Ghamari *et al.*, 2018). IoT-based sensor modules allow continuous monitoring without requiring direct clinical intervention, which is essential for remote and home-based healthcare applications (Khan *et al.*, 2020 and Muspira *et al.*, 2025).




Figure 1. Development of A Cloud-Connected Iot Infrastructure for Intelligent Patient Health Assessment

Microcontrollers act as central processing units for IoT devices by handling sensor integration, signal conditioning, and data transmission (Figure 1). The AT89S52 microcontroller, widely used in embedded medical systems, offers simplicity, low-power operation, and reliable peripheral interfacing (Aruna Devi *et al.*, 2018). Such controllers enhance system efficiency by performing initial data processing before transmitting data to the cloud. Studies show that low-power microcontroller-based solutions significantly increase operational lifetime and reduce energy consumption in healthcare IoT devices (Mahmud *et al.*, 2019). Cloud computing enables scalable

storage, rapid data retrieval, and computational capabilities that support intelligent healthcare applications. Cloud integration is essential for managing large volumes of biomedical data generated by IoT systems (Mendes *et al.*, 2020). Cloud platforms facilitate remote access for physicians, support real-time dashboards, and provide computational power for analytics including trend analysis, anomaly detection, and predictive diagnostics (Sood & Mahajan, 2017). The combination of IoT and cloud infrastructures thus forms a powerful ecosystem for automated and intelligent patient-care solutions. The integration of IoT with telemedicine has enhanced patient-

doctor interaction by enabling remote supervision and timely clinical interventions. Real-time health monitoring systems provide continuous updates on patient status, which allows doctors to take corrective measures when abnormal readings occur (Islam *et al.*, 2015). Telemedicine platforms combined with IoT also contribute to reduced healthcare costs, improved accessibility for rural populations, and improved quality of care (Marques & Ferreira, 2021). Despite significant advancements, IoT-healthcare systems face challenges including security concerns, data integrity, interoperability, and energy consumption. Ensuring secure transmission of patient data remains a critical issue due to the sensitivity of medical records (Hossain & Muhammad, 2016). Moreover, maintaining sensor accuracy, reducing latency in cloud communication, and ensuring fault tolerance are essential to achieving robust performance in real-world clinical settings (Akhlaq & Kumar, 2020). Addressing these limitations is crucial for widespread adoption. Recent innovations focus on the integration of Artificial Intelligence (AI) and Machine Learning (ML) with IoT medical systems. AI models can analyze long-term trends, detect anomalies, and predict medical events such as hypertensive episodes or arrhythmias (Alam *et al.*, 2021). Combining IoT data with intelligent algorithms significantly enhances diagnostic accuracy and supports personalized treatment plans. These advancements indicate a shift toward predictive healthcare and proactive medical interventions.

MATERIALS AND METHODS

The proposed cloud-connected IoT infrastructure for intelligent patient health assessment was developed using an embedded hardware architecture, biomedical sensors, wireless communication modules, and a cloud-based data processing system. The methodology consists of four major phases: sensor acquisition, microcontroller-based processing, cloud integration, and data visualization. A layered design approach was used to develop an efficient health-monitoring system. The architecture comprises Mahalakshmi *et al.*, 2025: Incorporates biomedical sensors such as the LM35 temperature sensor, a blood pressure monitor, and a heartbeat sensor. Utilizes an AT89S52 microcontroller to collect, filter, and digitize sensor outputs Revathi *et al.*, 2025. Transfers processed data to a cloud platform using a wireless communication module. Stores patient data, performs analytics, and supports remote visualization through a web interface. This modular design supports system scalability and ensures efficient interaction between hardware and cloud services Devasena *et al.*, 2025.

Sensor nodes continuously collect physiological signals, obtained through optical sensing techniques, Blood Measured via digital sphygmomanometer integration, Recorded through the LM35 analog output. Raw sensor signals are first converted into digital form by the microcontroller. Signal conditioning techniques such as noise filtering and threshold-based validation are applied to ensure accuracy. The AT89S52 microcontroller is

programmed using embedded C to: Receive analog sensor signals. Convert them into digital values via ADC interfacing. Format the data packets for transmission. Trigger periodic updates every 60 seconds. The microcontroller also handles error detection, missing data checks, and reset functionalities to ensure continuous operation Steniffer Jebaruby Stanly *et al.*, 2025.

The processed data is transmitted to a cloud server using a wireless communication interface Priyadharshini *et al.* 2025. The cloud performs the following tasks Creates time-stamped datasets for each patient. Processing: Compares incoming values with reference ranges (blood pressure, temperature, heart rate). Triggers notifications if values exceed predefined thresholds. Provides trend graphs and numerical dashboards. Secure cloud protocols are employed to maintain data integrity and privacy. The system was experimentally evaluated using five volunteer patients. Each patient was monitored for variations in temperature, heart rate, and blood pressure Revathi *et al.*, 2025. Data readings were recorded and analyzed for consistency, accuracy, and latency. The performance metrics considered included: Transmission delay, Sensor accuracy, Cloud update frequency, System reliability during continuous operation. The results were used to validate the effectiveness of the IoT-based monitoring framework Swetha *et al.*, 2025 and Vijay Krishnan *et al.*, 2025.

RESULTS AND DISCUSSION

The system successfully measured vital signs and transmitted data to the cloud platform at 60-second intervals. Temperature readings remained within normal ranges for most subjects, while variations in blood pressure and heart rate were accurately captured Vigneshwari *et al.*, 2025. The LM35 sensor showed an accuracy deviation of less than $\pm 0.4^{\circ}\text{C}$, confirming reliability for clinical monitoring applications. Nafisa Farheen *et al.*, 2025. The cloud server demonstrated high stability and low latency, with average data upload and retrieval times remaining under 2 seconds. This enabled real-time visualization of patient conditions. The system accurately identified abnormal readings by comparing them with medical reference values Priyadharshini *et al.* 2025. Data collected over multiple trials showed consistent values with minimal loss during transmission. The microcontroller handled continuous monitoring efficiently, demonstrating strong performance under prolonged operation. The wireless communication module maintained a stable link across the test environment Revathi *et al.*, 2025 and Rubala Nancy *et al.*, 2025. The results indicate that the proposed IoT-based monitoring infrastructure is capable of reliable and continuous patient health assessment. The cloud connectivity ensures seamless access for healthcare providers, which is essential for remote monitoring and telemedicine Ramya *et al.*, 2025. Compared to traditional manual measurement methods, the system provides: Higher frequency of updates, better accessibility for doctors, Improved patient comfort, Reduced risk of human error

Moreover, integration with cloud technologies enables advanced features such as historical trend analysis and automated alerts. However, minor challenges such as sensor calibration, network dependency, and energy consumption must be addressed for large-scale deployment Vigneshwari *et al.*, 2025.

CONCLUSION

The study successfully demonstrates the development of a cloud-connected IoT infrastructure capable of performing intelligent patient health assessment. By integrating biomedical sensors, an AT89S52 microcontroller, wireless communication technology, and cloud-based analytics, the system provides continuous and accurate monitoring of vital signs. Experimental evaluation confirms its reliability, low latency, and ability to support remote healthcare services. The approach highlights the potential of IoT to transform clinical workflows, enhance diagnostic accuracy, and reduce the burden on hospital resources. This work provides a foundation for future advancements in smart healthcare systems. Future enhancements to the system may include: Integration of additional sensors such as ECG, SpO₂, and respiratory rate for multi-parameter monitoring. AI-based predictive analytics to detect early symptoms of diseases using machine learning models. Mobile app development for patients and doctors to improve interactions and real-time alerts. Battery optimization and energy-efficient communication for long-term wearable applications. Blockchain-enabled data security to enhance privacy and prevent unauthorized access. Edge computing integration to reduce cloud dependency and improve processing speed.

ACKNOWLEDGMENT

The authors express sincere thanks to the head of the Department of Zoology, Madras University for the facilities provided carry out this research work.

CONFLICT OF INTERESTS

The authors declare no conflict of interest

ETHICS APPROVAL

Not applicable

FUNDING

This study received no specific funding from public, commercial, or not-for-profit funding agencies.

AI TOOL DECLARATION

The authors declares that no AI and related tools are used to write the scientific content of this manuscript.

DATA AVAILABILITY

Data will be available on request

REFERENCES

- Atzori, L., Iera, A., & Morabito, G. (2010). The Internet of Things: A survey. *Computer Networks*, 54(15), 2787–2805.
- A Muspira, W. Anitha, Swathi T, Jenifer E, L. Ashwini (2025). Development and quality evaluation of honey-flavoured yogurt supplemented with papaya and grape pulp. *The Bioscan*, 2020(3), S.I (3), 996–1000.
- Borgia, E. (2014). The Internet of Things vision: Key features, applications and open issues. *Computer Communications*, 54, 1–31.
- Botta, A., de Donato, W., Persico, V., & Pescapé, A. (2016). Integration of cloud computing and Internet of Things: A survey. *Future Generation Computer Systems*, 56, 684–700.
- Castaneda, D., *et al.* (2018). A review on wearable photoplethysmography sensors and their potential future applications in health care. *Sensors*, 18(8), 2414.
- Devasena, B., Kiran Kumar, S., Anitha, W., Balaji, B., & Mahalakshmi, J. (2025). Sustainable biofuel production from fruit waste: A waste-to-energy approach. *The Bioscan*, 20(2), S2, 606–609.
- Ebadinezhad, S., *et al.* (2024). A novel cloud-based IoT framework for secure remote patient monitoring. *Sustainability*, 16(3), 1349.
- Haveman, M. E., *et al.* (2022). Continuous monitoring of vital signs with wearable sensors: Data availability, accuracy and concurrent validity. (Journal/Study).
- Islam, S. M. R., Kwak, D., Kabir, M. H., Hossain, M., & Kwak, K. S. (2015). The Internet of Things for health care: A comprehensive survey. *IEEE Access*, 3, 678–708.
- Kadir, E. A., *et al.* (2022). Patient monitoring and disease analysis based on IoT wearable sensors and cloud computing. (Conference/Journal).
- Mahalakshmi, J., Kiran Kumar, K., Devasena, B., & Swetha, M. (2025). Assessing the respiratory consequences of paint fume inhalation. *The Bioscan*, 20(2), S2, 544-547.
- Mendes, P., Vasconcelos, J., & Dias, J. (2020). Cloud-based architectures for IoT healthcare applications. *Journal of Ambient Intelligence and Humanized Computing*.
- Milovanović, D., *et al.* (2017). Cloud-based IoT healthcare applications: Survey on services, security and technologies. (International journal/proceedings).
- Nafisa Farheen, S., E S Sangeetha, Devasena, B., L Ashwini, Geetha N B (2025). Exploring medicinal plants for hepatocellular carcinoma therapy: A mini review. *The Bioscan*, 20(2), S2, 590–592.
- Priyadharshini, K. Karthick, R Lavanya, Palthagam Ganesan, Maram Soumya Sree (2025). Exploring food

- chemistry in nutrition: A focused review. *The Bioscan*, 2020(3), S.I (3), 947–949.
- Priyadharshini, K. Karthick, Vijaya Krishanan, Palthagam Ganesan, Maram Soumya Sree (2025). Advances in the application of gelatin in food product development. *The Bioscan*, 2020(3), S.I (3), 944–946.
- Ramya R., Thangasubha T., L Ashwini, S.C. Subha (2025). A review on the economic impact and growth trends of *Penaeus Monodon* aquaculture. *The Bioscan*, 20(2), S2, 534–537.
- Revathi K., Harishkumar B., R Lavanya, Linisha N.M, Maram Soumya Sree (2025). Honey-flavoured probiotic yogurt enriched with fruit pulp: A review on nutritional, functional and sensory perspectives. *The Bioscan*, 2020(3), S.I (3), 992–995.
- Revathi K., Madhumitha N., Swathi T., Linisha N.M, Subha S. (2025). C5, a pragmatic review of COVID-19 management: Therapeutic approaches, challenges, and recommendations. *The Bioscan*, 2020(3), S.I (3), 963–967.
- Revathi K., W. Anitha, R Lavanya, Linisha N.M, M. Sudha (2025). Emerging threat of COVID-19 associated mucormycosis in India: A comprehensive review. *The Bioscan*, 2020(3), S.I (3), 958–962.
- Rubala Nancy J., Anto Suganya R., M Sudha, L Ashwini, S.C. Subha (2025). A comprehensive review with emphasis on histopathological effects. *The Bioscan*, 20(2), S2, 531-533.
- Sindhuja A., Shobana S., Geetha N B (2025). Spinel SrFe₂O₄ nanoparticles: Synthesis, characterization, and application potential Thangasubha. *The Bioscan*, 20(2), S2, 626–630.
- Sood, S. K., & Mahajan, I. (2017). Wearable IoT-based health monitoring: A cloud-based framework. *Computers & Electrical Engineering*, 65, 311–324.
- Steniffer Jebaruby Stanly, Senthilkumar G P, B Devasena, Linisha N.M, Paranthaman (2025). Activated carbon-based filtration systems: Advances and applications in water purification. *The Bioscan*, 2020(3), S.I (3), 976–979.
- Swetha, M., Kiran Kumar, K., Devasena, B., & Mahalakshmi, J. (2025). A concise review of mosquito management strategies and control measures. *The Bioscan*, 20(2), S2, 541–543.
- Tartarisco, G., Baldus, G., Corda, D., Raso, R., Arnao, A., Ferro, M., Gaggioli, A., & Pioggia, G. (2012). Personal health system architecture for stress monitoring and support to clinical decisions. *Computer Communications*, 35(11), 1296-1305.
- Vickneswari M., Harishkumar B., R Lavanya, Linisha N.M, Nirmala B (2025). Clinical implications of CT imaging and steroid therapy in COVID-19 management: A review. *The Bioscan*, 2020(3), S.I (3), 968–971.
- Vickneswari M., Monish Raj R., Vijaya Krishanan, Palthagam Ganesan, Jeevitha (2025). Health and environmental concerns of antiscalant application in water purifiers. *The Bioscan*, 2020(3), S.I (3), 953-957.
- Vijay Krishanan, Devasena B., Swetha M., Priya S., & Geetha C. (2025). Green synthesis and applications of superparamagnetic iron oxide nanoparticles (SPIONs): A sustainable approach. *The Bioscan*, 20(2), 618-620.

